DECLASSIFIED

UCRL 416

~~Copy 83~~

CONFIDENTIAL

UNIVERSITY OF CALIFORNIA

Radiation Laboratory

TWO-WEEK LOAN COPY

This is a Library Circulating Copy which may be borrowed for two weeks. For a personal retention copy, call Tech. Info. Division, Ext. 5545

BERKELEY, CALIFORNIA

UCRL-416 c.2

Special Review of
Declassified Reports

Authorized by USDOE JK Bratton

Unclassified TWX P182206Z May 79

REPORT PROPERLY DECLASSIFIED

By JB Stuart 8-20-79

Authorized Derivative Classifier Date

DISCLAIMER

This document was prepared as an account of work sponsored by the United States Government. While this document is believed to contain correct information, neither the United States Government nor any agency thereof, nor the Regents of the University of California, nor any of their employees, makes any warranty, express or implied, or assumes any legal responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by its trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof, or the Regents of the University of California. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof or the Regents of the University of California.

UNIVERSITY OF CALIFORNIA
Radiation Laboratory

Cover Sheet
Do not remove

INDEX NO. UCRL 416
This document contains 7 pages
This is copy 83 of 92 Series A

DECLASSIFIED

Issued to Info Division

~~CONFIDENTIAL~~

Classification

Each person who receives this document must sign the cover sheet in the space below

Route to	Noted by	Date	Route to	Noted by	Date

CONFIDENTIAL

UCRL 416
Physics-General Distribution

DECLASSIFIED

UNIVERSITY OF CALIFORNIA

Radiation Laboratory

Contract No. W-7405-eng-48

CLASSIFICATION CANCELLED
BY AUTHORITY OF THE DECLASSIFICATION
BRANCH USAEC per TID 1113
BY B Frorkett 3-31-56
SIGNATURE OF THE PERSON MAKING THE CHANGE DATE

SUMMARY OF THE RESEARCH PROGRESS MEETING OF JULY 28, 1949

H. P. Kramer

August 31, 1949

Special Review of Declassified Reports
Authorized by USDOE JK Bratton
Unclassified TWX P182206Z May 79

REPORT PROPERLY DECLASSIFIED
AB Stuart 8-20-79
Authorized Derivative Classifier Date
Marilyn Pennington 1-22-80
By Date

CAUTION

This document contains information affecting the National Defense of the United States. Its transmission or the disclosure of its contents in any manner to an unauthorized person is prohibited and may result in severe criminal penalties under applicable Federal laws.

Berkeley, California

Standard Distribution	Copy Numbers
Argonne National Laboratory	1-12
Armed Forces Special Weapons Project	13
Atomic Energy Commission, Washington	14-15
Battelle Memorial Institute	16
Brookhaven National Laboratory	17-20
Bureau of Ships	21
Carbide & Carbon Chemicals Corporation (K-25 Plant)	22-25
Carbide & Carbon Chemicals Corporation (Y-12 Plant)	26-29
Columbia University (Dunning)	30
General Electric Company, Richland	31-34
Hanford Operations Office	35
Iowa State College	36
Knolls Atomic Power Laboratory	37-40
Los Alamos	41-43
Mound Laboratory	44-46
National Advisory Committee for Aeronautics	47
National Bureau of Standards	48-49
Naval Radiological Defense Laboratory	50
NEPA Project	51
New Brunswick Laboratory	52
New York Operations Office	53-54
North American Aviation, Inc.	55
Oak Ridge National Laboratory	56-63
Patent Branch, Washington	64
Sandia Laboratory	65
Technical Information Branch, ORE	66-80
UCLA Medical Research Laboratory (Warren)	81
University of California Radiation Laboratory	82-86
University of Rochester	87-88
Westinghouse Electric Corporation	89-92
	92

INFORMATION DIVISION
Radiation Laboratory
Univ. of California
Berkeley, California

SUMMARY OF THE RESEARCH PROGRESS MEETING OF JULY 28, 1949

H. P. Kramer

August 31, 1949

Scintillation Counters for General Research Use. L. Wouters.

A number of scintillation counters have been designed and produced and are available for general research. Two mountings are available. One consists of a single crystal and photomultiplier tube and the other one contains two photomultiplier tubes and is to be used with a single crystal in detecting radiation that produces low ionization in the crystal.

Each mounting is equipped with a switch that will allow the operator to use the instrument in connection with a linear amplifier or else with a linear amplifier in conjunction with a feed-back loop to produce a square pulse.

The counters are expected to be capable of registering about 3000 counts/sec. in coincidence and about 10^5 counts/sec. when used separately. The resolving time of the circuit is about .25 microseconds. It depends on the time constant of the RC loop, .2 microseconds, which cannot be reduced further without electron accumulation at the counting end of the circuit.

In the present design, electronic and mechanical counting devices are used in preference to an oscilloscope. In the design of a proper mixing unit for coincidence counts the discrimination between almost coincident and coincident counts poses a delicate problem. In an improperly designed mixing unit pulses that are separated in time may show as a coincidence because of a random distribution of pulse heights and widths. A partial solution of this difficulty has been achieved in the present apparatus. All pulses are of the same height, 30 volts, and last for .15 microseconds. Pulses that overlap more than .05 microseconds will be counted as coincidences.

The present equipment has been designed for use with anthracene crystals. Anthracene was chosen in preference over the faster transtilbine because it produces bigger pulses. Diphenyl acetylene, a material that has newly been suggested for scintillation counters, may combine the advantages of both anthracene and transtilbine.

The Study of U^{231}. W. Crane.

U^{231} decays both by K-capture and α-emission. As a step in the general development of regularities among heavy α-emitters it was decided to make a detailed examination of the mode of decay of U^{231}. The isotope was produced by the bombardment of thorium with 150 Mev helium ions. By milking the α-daughter Th^{227} and the K-capture daughter Pa^{231} it was possible to determine the half-life for α-decay as being 208 years and to fix the K-capture half-life at 4.2 days.

The following equations embody the assumed mechanism of K-capture decay and suggest alternative experimental procedures:

$$C_K \cdot 1/\varepsilon_K \cdot 1/\rho_K = N \qquad (1)$$

$$C_L \cdot 1/\varepsilon_L \cdot 1/\rho_L \cdot 1/\mu_\alpha + 1 - \rho_K = N \qquad (2)$$

$$Q_{K-L} \cdot 1/\varepsilon_L \cdot 1/\rho_L \cdot 1/\varepsilon_K \cdot 1/\rho_K \cdot \frac{\mu_\alpha + 1 - \rho_K}{\mu_\alpha^2} = N \qquad (3)$$

$$Q_{L-L} \cdot 1/\varepsilon_L^2 \cdot 1/\rho_L^2 \cdot 1/\mu_\alpha^2 \, (1 - \rho_K + \mu_\alpha)^2 = N \qquad (4)$$

In these equations, N represents the number of K-capture disintegrations per minute, C the number of counts per minute, Q the number of coincidences per minute, ρ the fraction of x-rays produced that remains unconverted, and ε the external counting efficiency which is the fraction of emitted x-rays of one kind that is actually detected by the counters. The external counting efficiency ε therefore involves both geometrical and electronic considerations. The subscripts refer to the place of origin of the x-rays. The index K, for example,

shows that the quantity under consideration pertains to x-rays that are emitted when a vacancy in the K shell is filled by an electron from a higher level. A certain fraction μ_α of all K x-rays produced arises when the vacancy in the K shell is filled by an electron from the L shell. Such K x-rays are called K_α x-rays.

Equation (1) expresses the assumption that every disintegration proceeds by the nuclear capture of an electron from the K shell. Every disintegration thus leaves a vacancy in the K shell which must be filled by the drop of an electron from a higher level and therefore must be accompanied by the production of a K-x-ray. The production of a K_α x-ray is accompanied by the production of an L x-ray since a vacancy is left in the L shell when one of its electrons serves to replenish the K shell. The other source of L x-rays is the internal conversion of K x-rays in the L shell. A K x-ray yields its energy to an electron in the L shell and thereby causes its emission as an Auger electron from the atom. In this way a vacancy arises in the L shell. When this vacancy is filled from a higher level an L x-ray is produced. These notions are expressed in equation (2). It is assumed throughout this work that all K x-rays which are converted are converted in the L shell and thus serve to produce L x-rays. By making use of equations (1) and (2), equation (3) may also be expressed in the form:

$$C_K \cdot C_L/Q_{K-L} \cdot \mu_\alpha^2/(\mu_\alpha+1-\rho_K)^2 = N$$

Finally, equation (4) expresses the supposition the L-L coincidences arise by the simultaneous detection of L x-rays that are produced in conjunction with K_α x-rays and L x-rays whose production is incident on the conversion of K x-rays in the L shell.

An attempt has been made to determine the fraction of all disintegrations that culminate in the ejection of an Auger electron by a photographic technique. A drop of solution containing U^{230} and U^{231} was placed on an electron sensitive

photographic plate and the number of electron tracks from U^{231} conversions and the number of α-stars from U^{230} disintegrations were counted. Some difficulty was caused by the high electron background. Using U^{230} as a time standard, the fraction of Auger electron emissions among the total number of disintegrations was found to be .08 on the basis of the disintegration rate that was determined from the time wise yield of Pa^{231}.

The number of coincidences was counted by two tubes separated by an angle of 180°. The x-rays were analyzed by various foils placed between a mica sheet permeated with active solution and the counter windows. In coincidence counting, Auger electrons also were stopped by the foils and counter windows. A typical arrangement is sketched in Fig. 1. The efficiency for counting L x-rays was considerably higher than that for counting K x-rays: $\varepsilon_L = 0.5$ percent and $\varepsilon_K = 0.2$ percent. For this reason the L-L coincidence rate was about 10 times as large as the K-L coincidence rate.

Seventy-eight percent of all disintegrations produce K_α x-rays and therefore also L x-rays, that is, $\mu_\alpha = .78$. Eight percent of all disintegrations give rise to Auger conversion electrons and hence presumably to L x-rays. Therefore, 86 percent of all disintegrations are expected to be accompanied by the emission of L x-rays. Only 51 percent of the disintegrations are however observed to be accompanied by L x-ray emission. It is therefore assumed that 42 percent of the L x-rays are internally converted.

At present, work is continuing in view of the fact that when the observed value for Q_{L-L} is substituted in equation (4) one arrives at a value for N that is four times as large as that observed in milking experiments. In order to explain this discrepancy, it may be necessary to postulate a γ-ray that is completely internally converted primarily in the L shell.

Information Division
scb/9-20-49



Fig. 1

DECLASSIFIED

CONFIDENTIAL