

UCRL 743
cy. 2

UNCLASSIFIED

UNIVERSITY OF
CALIFORNIA

*Radiation
Laboratory*

TWO-WEEK LOAN COPY

*This is a Library Circulating Copy
which may be borrowed for two weeks.
For a personal retention copy, call
Tech. Info. Division, Ext. 5545*

BERKELEY, CALIFORNIA

DISCLAIMER

This document was prepared as an account of work sponsored by the United States Government. While this document is believed to contain correct information, neither the United States Government nor any agency thereof, nor the Regents of the University of California, nor any of their employees, makes any warranty, express or implied, or assumes any legal responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by its trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof, or the Regents of the University of California. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof or the Regents of the University of California.

UNCLASSIFIED

UCRL-743

Unclassified Distribution

COPY 2

UNIVERSITY OF CALIFORNIA

Radiation Laboratory

Contract No. W-7405-eng-48

DETECTION OF TRITIATED COMPOUNDS IN PAPER CHROMATOGRAPHY

Irving Gray, Saburo Ikeda, A. A. Benson, and David Kritchevsky

June 5, 1950

Berkeley, California

INSTALLATION:	No. of Copies
Argonne National Laboratory	8
Armed Forces Special Weapons Project	1
Atomic Energy Commission, Washington	2
Battelle Memorial Institute	1
Brush Beryllium Company	1
Brookhaven National Laboratory	8
Bureau of Medicine and Surgery	1
Bureau of Ships	1
Carbide and Carbon Chemicals Div., Union Carbide and Carbon Chemicals Corp. (K-25 Plant)	4
Carbide and Carbon Chemicals Div., Union Carbide and Carbon Chemicals Corp. (Y-12 Plant)	4
Chicago Operations Office	1
Cleveland Area Office, AEC	1
Columbia University (J. R. Dunning)	2
Columbia University (G. Failla)	1
Dow Chemical Company	1
H. K. Ferguson Company	1
General Electric Company, Richland	3
Harshaw Chemical Corporation	1
Idaho Operations Office	1
Iowa State College	2
Kansas City Operations Branch	1
Kellex Corporation	2
Knolls Atomic Power Laboratory	4
Los Alamos Scientific Laboratory	3
Mallinckrodt Chemical Works	1
Massachusetts Institute of Technology (A. Gaudin)	1
Massachusetts Institute of Technology (A. R. Kaufmann)	1
Mound Laboratory	3
National Advisory Committee for Aeronautics	2
National Bureau of Standards	2
Naval Radiological Defense Laboratory	2
New Brunswick Laboratory	1
New York Operations Office	5
North American Aviation, Inc.	1
Oak Ridge National Laboratory	8
Patent Branch, Washington	1
Rand Corporation	1
Sandia Laboratory	1
Santa Fe Operations Office	1
Sylvania Electric Products, Inc.	1
Technical Information Division, Oak Ridge	15
USAF, Air Surgeon (R. H. Blount)	1
USAF, Director of Armament (C. I. Browne)	1
USAF, Director of Plans and Operations (R. L. Applegate)	1
USAF, Director of Research and Development (F. W. Bruner, and R. J. Mason)	2
USAF, Eglin Air Force Base (A. C. Field)	1

INSTALLATION:	No. of Copies
USAF, Kirtland Air Force Base (M. F. Cooper)	1
USAF, Maxwell Air Force Base (F. N. Moyers)	1
USAF, NEPA Office	2
USAF, Office of Atomic Energy (A. A. Fickel, H. C. Donnelly)	2
USAF, Offutt Air Force Base (H. R. Sullivan, Jr.)	1
USAF, Wright-Patterson Air Force Base (Rodney Nudenberg)	1
U. S. Army, Atomic Energy Branch (A. W. Betts)	1
U. S. Army, Army Field Forces (James Kerr)	1
U. S. Army, Commanding General, Chemical Corps Technical Command (J. A. MacLaughlin thru Mrs. G. Benjamin)	1
U. S. Army, Chief of Ordnance (A. R. Del Campo)	1
U. S. Army, Commanding Officer Watertown Arsenal (C. H. Deitrick)	1
U. S. Army, Director of Operations Research (Ellis Johnson)	1
U. S. Army, Office of Engineers (Allen O'Leary)	1
U. S. Army, Office of the Chief Signal Officer (Curtis T. Clayton thru G. C. Hunt)	1
U. S. Army, Office of the Surgeon General (W. S. Stone)	1
U. S. Geological Survey (T. B. Nolan)	1
U. S. Public Health Service	1
University of California at Los Angeles	1
University of California Radiation Laboratory	5
University of Rochester	2
University of Washington	1
Western Reserve University	2
Westinghouse Electric Company	4
University of Rochester (R. E. Marshak)	1
California Institute of Technology (Dr. Robert F. Bacher)	1
Total	144

Information Division
Radiation Laboratory
University of California
Berkeley, California

DETECTION OF TRITIATED COMPOUNDS IN PAPER CHROMATOGRAPHY

Irving Gray⁽¹⁾, Saburo Ikeda, A. A. Benson and David Kritchevsky

Radiation Laboratory and Department of Chemistry,
University of California, Berkeley, Calif. ⁽²⁾

ABSTRACT

June 5, 1950

A windowless tube has been developed which can be used to detect tritium radiations on paper, glass and alumina. The detection of tritiated compounds in paper chromatography is greatly simplified by the use of this device.

For publication in Review of Scientific Instruments

-
- (1) Major, Medical Service Corps, U. S. Army. Present address: Brooke Army Medical Center, Fort Sam Houston, Texas.
- (2) The work described in this paper was sponsored by the Atomic Energy Commission.
-

DETECTION OF TRITIATED COMPOUNDS IN PAPER CHROMATOGRAPHY

Irving Gray,⁽¹⁾ Saburo Ikeda, A. A. Benson and David Kritchevsky

Radiation Laboratory and Department of Chemistry,
University of California, Berkeley⁽²⁾

Location of tritiated compounds in paper chromatography has posed a problem because of the difficulties involved in the detection of the active spots. Because of the low energy of its emitted β -rays, the assay of tritium has usually been done in the gaseous phase; this procedure does not lend itself to the location of tritiated compounds on paper. A possible method, which has been tried in this laboratory, involves cutting of the paper into squares and counting each square in a windowless counter such as the "Nucleometer." This method is disadvantageous since it renders the paper useless for further work. A counting tube has been developed which will measure the radiation from the spots in question easily and accurately.

The counting apparatus consists of a windowless Scott tube⁽³⁾, a gas flow regulator and a source of helium saturated with alcohol at 0° C. The

-
- (1) Major, Medical Service Corps, U. S. Army. Present address: Brooke Army Medical Center, Fort Sam Houston, Texas.
 - (2) The work described in this paper was sponsored by the Atomic Energy Commission.
 - (3) A bell-shaped glass tube, lined internally with an electrode consisting of copper sulfide coated copper foil, the open end which has a diameter of 7/8 inch covered with a thin mica window, ca. 1.5 mg./cm² and having a nichrome electrode extending through the center of the tube. The center electrode is tipped with a glass bead. This tube was developed by Dr. Kenneth G. Scott of this laboratory.
-

tube is wrapped in lead foil to minimize background radiation and to add ballast, so that a constant resistance to gas flow through the paper is maintained.

A rate of gas flow of 300 ml./min. was maintained. This rate insured a reproducible counting rate. At this pressure the system showed the following characteristics:

Starting voltage	1125 volts
Threshold voltage	1200 volts
Plateau	200 volts
Optimum operating voltage	1350 volts

For scanning, the tube is allowed to rest on the paper and is slowly moved over the paper in a definite pattern. The location of activity is recorded on a scaler. Where a particular area is to be counted, the tube is flushed for at least thirty seconds, after which the scaler is turned on and the activity determined.

This apparatus has been used to scan and count papers holding either C^{14} or tritium containing compounds. Table I summarizes the results obtained using various types of counters. For counting in the "Nucleometer", papers were cut into appropriately sized pieces.

Table I

Comparison of the values obtained by counting the same sample by various counting devices. All data given as counts per minute x 64.

<u>Sample</u>	<u>Scott Tube</u>	<u>Windowless Tube</u>	<u>Nucleometer</u>
* T (on paper)	a	191;193;193	173;201;133 ^b
T (on glass plate)	a	118;119;115	210;191;172 ^b
T (on aluminum plate)	a	47;47;48	93;94;93
** C (on paper)	60;64;66	95;93;97	164;168;173

* Denotes tritium-containing compound (tritiated cholesterol)

** Denotes C^{14} -containing compound (stearic- $l-C^{14}$)

a No measureable activity

b Very erratic and unsteady counting

Once the active spots have been located on the paper, it is desirable to make a radiogram of the paper in order to have an accurate outline of the active area (4). Here again, the low energy of tritium radiation (0.015 mev) and the high self-absorption indicated the necessity of altering the usual procedure. We have found, however, that a satisfactory image of an area containing a tritiated compound is obtained on Eastman "No-Screen" X-ray film in twenty-four hours, if the area has an activity of 15,000 counts/min/cm². as recorded by the windowless tube. (C^{14} , 0.16 mev, requires a recorded activity of about

(4) A. A. Benson, et. al., J. Am. Chem. Soc., 72, 1710 (1950).

200 counts/min/cm² in order to give a satisfactory image in twenty-four hours.

Acknowledgment: The authors wish to acknowledge the kind interest shown by Prof. Melvin Calvin and Dr. Bert M. Tolbert during the course of this investigation.

SUMMARY

A windowless tube has been developed which can be used to detect tritium radiations on paper, glass and alumina. The detection of tritiated compounds in paper chromatography is greatly simplified by the use of this device.