

UNIVERSITY OF  
CALIFORNIA

*Ernest O. Lawrence*

*Radiation  
Laboratory*

TWO-WEEK LOAN COPY

*This is a Library Circulating Copy  
which may be borrowed for two weeks.  
For a personal retention copy, call  
Tech. Info. Division, Ext. 5545*

BERKELEY, CALIFORNIA

## **DISCLAIMER**

This document was prepared as an account of work sponsored by the United States Government. While this document is believed to contain correct information, neither the United States Government nor any agency thereof, nor the Regents of the University of California, nor any of their employees, makes any warranty, express or implied, or assumes any legal responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by its trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof, or the Regents of the University of California. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof or the Regents of the University of California.

UCRL-9272 Rev.  
Limited distribution

UNIVERSITY OF CALIFORNIA  
Lawrence Radiation Laboratory  
Berkeley, California  
Contract No. W-7405-eng-48

A TRANSISTOR COINCIDENCE-DISCRIMINATOR CIRCUIT  
FOR  $10^{-8}$ -SECOND PULSES

Stanley C. Baker

March 6, 1961

A TRANSISTOR COINCIDENCE-DISCRIMINATOR CIRCUIT  
FOR  $10^{-8}$ -SECOND PULSES

Stanley C. Baker

Lawrence Radiation Laboratory  
University of California  
Berkeley, California

March 6, 1961

ABSTRACT

A transistor coincidence-discriminator circuit of better than  $10^{-8}$ -sec resolution and 0.25-v sensitivity has been designed for use where large numbers of such circuits are required in a single system. The experience gained in operating 168 units in a physics experiment, and the methods developed for testing the sensitivity and resolution of the circuits during both initial checkout and operation in the experiment show the approach used to circuit design to be highly feasible.

A TRANSISTOR COINCIDENCE-DISCRIMINATOR CIRCUIT  
FOR  $10^{-8}$  - SECOND PULSES\*

Stanley C. Baker<sup>†</sup>

Lawrence Radiation Laboratory  
University of California  
Berkeley, California

March 6, 1961

INTRODUCTION

For several years, multiple-counter experiments have been performed at this Laboratory using two-fold coincidence circuits in which each channel receives pulses from a scintillation counter and a common gate pulse. The gate pulse, usually generated by a counter system, indicates the passage of a particle with a particular time of flight or momentum. Thus all coincidence channels are sampled for the gate period whenever the desired event is detected.

A pulse-amplitude discriminator follows the coincidence circuit to accurately determine which events produce signals above a predetermined threshold level and should be recorded. To relax the bandwidth requirements of the system after a coincidence has been detected, a memory circuit is employed to preserve the response of each event until further logical operations are completed.

Multiple-diode coincidence circuits whose outputs are appropriately delayed and mixed for display on an oscilloscope have been used in numerous experiments.<sup>1</sup> A permanent record is then made on photographic film for later analysis.

---

\* This work was done under the auspices of the U. S. Atomic Energy Commission.

<sup>†</sup> Now at Cubic Corporation, San Diego, California.

When a 168-channel system was proposed, it became evident that the experimental data should be acquired and preserved in a form suitable for immediate entry into a computer.<sup>2</sup> The coincidence-discriminator circuit described here was developed to accept pulses directly from multiplier phototubes without preamplification and to store the coincidence patterns electrically instead of photographically.

Development specifications for the unit are the following: The minimum input-amplitude sensitivity must be 0.25 v into an impedance of 125 ohms, but 10 v signals must not cause adverse effects. The memory must accept information at a rate of at least a 100 kC; however, 10 v noncoinciding pulses at either the phototube or gate inputs occurring with a frequency of 10 Mc must not cause accidental output signals. The time resolution of the coincidence circuit should be 5 nsec or less. The unit must operate satisfactorily up to an ambient temperature of 55°C. Since more than one coincidence circuit would be operated from each phototube signal, a means of operating units in parallel should be provided. The output of the coincidence-circuit portion of the unit must be monitored without affecting the operation of the remainder of the unit.

## CIRCUIT DESCRIPTION

At the time of development, the most economically feasible transistors with a sufficient gain-bandwidth product (i. e.  $f_T > 100$  Mc) were PNP types. A diffused base, mesa unit, type 2N1143, was selected for use in the fast circuits. It has a minimum gain-bandwidth product ( $f_T$ ) of 250 Mc at a collector potential of 10 v at 10 ma. Using PNP transistors meant that coincidence-circuit limiters had to receive positive base signals to cut off their collector currents.

A block diagram of the coincidence discriminator that was developed is shown in Fig. 1, and the schematic circuit diagram in Fig. 2.

The minimum acceptable input signal turned out to be 100 mv, which gave an adequate margin in sensitivity over the specified value of 250 mv. A phase-splitting stage, transistor Q-1, was used to provide isolation between the portion of the signal fed to the coincidence circuit and that part which might be used to drive another unit or a phototube-signal monitor. For proper operation, the monitor jack is terminated in an impedance of 125 ohms. The input stage Q-1 inverts negative phototube signals in order to apply positive pulses to the coincidence limiters.

Since the pulse length of phototube signals may vary considerably, a 9-nsec, 330-ohm, coaxial clipping stub was employed in the collector circuit of Q-1 to limit this length to 18 nsec. The collector load resistance in parallel with the input impedance of transistor Q-2 provides a termination for the clipping stub. The use of Q-2 as an emitter follower provided a better impedance match for the clipping line than that afforded by feeding the clipped signal directly to limiter Q-3.

The gate pulse, which is applied in common to each coincidence unit, is externally shaped to a specified height and width. Careful design was required in the gate distribution system to drive two groups of 84 coincidence units with a minimum delay and pulse lengthening. Distribution was made by first splitting the gate into three channels driven by a complementary PNP - NPN emitter follower. Each of these channels is then multiplied into six more channels in another emitter-follower unit. These channels are finally separated into five more to provide the gate signals to drive each coincidence unit.

A schematic circuit diagram of the last 5-channel gate driver is shown in Fig. 3. A type 2N1143 emitter follower drives five grounded-emitter amplifiers in parallel. The gain is 0.5 for each channel loaded with 125 ohms. It provides a positive output signal 0.5 v in amplitude and 20 nsec long.

A Garwin coincidence circuit was selected.<sup>3</sup> Diode D-2 is the one normally associated with the Garwin circuit; it draws 55% of the combined collector currents of limiters Q-3 and Q-4. Both transistors Q-3 and Q-4 must be cut off before D-2 unclamps and allows an output signal. The output is directed to an amplitude-discriminating diode D-3. The coincidence ratio (doubles-to-singles) is approximately 10/1.

Several other coincidence circuits were investigated, but this one proved to be best suited to our needs. In other circuits degenerating resistors in the emitters of the limiters caused the input impedance to be much higher and allowed a better termination for a transmission line than could be obtained with the circuit actually employed. This, however, increased the pulse height needed for limiting. Decreased sensitivity was undesirable, while matching impedances was not a great problem, and so unbypassed emitter resistors were not used.

To obtain more output signal, the circuit of Fig. 4 was tried, in an attempt to get a good coincidence ratio without sacrificing half the output current, as is done with the Garwin clamp diode. When one transistor in Fig. 4 cuts off, the other side assumes the whole current; when both sides are cut off, the whole bias current is available as a pulse in the output. This worked well and provided the same resolution time as the Garwin circuit, but again the loss of sensitivity was too disadvantageous. This loss is, of course, due to the presence of the emitter resistor.

Amplifier Q-5 and blocking oscillator Q-6 serve to convert signals with an amplitude greater than the threshold bias on diode D-3 into a 4-v, 0.8- $\mu$ sec pulse. A blocking oscillator was chosen to generate this pulse, because it needs only one transistor and possesses a very sensitive triggering action. When biased beyond cut off, it exhibits good threshold stability. The pulse produced has a very stable width, and operation is good over a wide range of temperature.

To make sure that transients produced by the blocking oscillator did not feed back into the sensitive portion of the circuit, the power supplied to the trigger circuit had to be well-bypassed.

The capacitance loading by the memory flip-flop Q-7 and Q-8 reduces the input sensitivity of the blocking oscillator. Diode D-4 provides some isolation for the blocking oscillator during its regenerative period.

Memory for coincident events is provided by an Eccles-Jordan bistable circuit. Two output connections are buffered by emitter followers Q-9 and Q-10. The flip-flop is reset by a positive pulse applied to diode D-6 after the necessary logic functions have been completed by the rest of the readout system.

Each unit was constructed on a glass-epoxy-laminate, printed-circuit board measuring 3 by 10 1/2 in. A photograph of this unit is shown in Fig. 5. Ten coincidence-discriminator units and two gate-circuit driver boards were plugged into

a frame measuring 3 1/2-in. high by 19 in. wide. Phototube-signal and coincidence-monitor connections were made to a front panel on each unit; all other connections were made through connections at the rear of the board.

During initial line-up as well as during the running of multi-channel experiments it was desired to check the operation of the unit--especially the threshold sensitivity.<sup>4</sup> A test signal (introduced in parallel with the phototube input and in time coincidence with a test gate pulse) provided a means of making this measurement. A pulser was constructed to allow test signals to be introduced in this manner by either a single pulse or at a 800 cps rate. The pulses were shaped to resemble the actual phototube signals.

## OPERATION

The amplitude threshold of the coincidence-discriminator circuit was found to be a strong function of the dc supply voltages. The -5-v bias potential had to be kept within 0.1% to keep the threshold variation within 1%. The sensitivity to the -15-volt supply line was not as great; it had to be maintained within 1% to hold the threshold within 1%. These voltages were monitored by a 2% full-scale recorder during operation.

The drift of the threshold level with time was quite acceptable. At the Bevatron without temperature control during a normal 8-hr period, four or five of the 168 units would drive outside of a  $\pm 5\%$  range. This was measured by the checking system previously mentioned.<sup>4</sup> It was a simple matter to periodically readjust the threshold levels of these units that had drifted outside the prescribed range.

During the early part of the experiment, approximately 12 transistors were burned out in the input amplifier Q-1. It was found that switching on the high voltage for the phototubes caused a large positive transient on the signal line. A protective diode, D-1, is now placed across the input circuit.

### ACKNOWLEDGMENTS


The circuit and package design was done in cooperation with H. G. Jackson. Stuart Wright checked and maintained each circuit both before and during operation.

REFERENCES

1. See, for example, Fig. 5 of C. A. Coombes, B. Cork, W. Galbraith, G. R. Lambertson, and W. A. Wenzel, Phys. Rev. 112, 1303 (1958).
2. S. C. Baker, F. A. Kirsten, D. A. Mack, and C. E. Wiegand, An Electronic Data-Aquisition System for Use with a Complex Nuclear Physics Experiment, UCRL-9271, December 1960; submitted to Nuclear Instruments and Methods.
3. R. L. Garwin, Rev. Sci. Instr. 24, 618 (1953).
4. S. C. Baker, Q. A. Kerns, F. A. Kirsten, D. A. Mack, and T. A. Nunamaker, Test Routines and Monitoring, UCRL-9273, December 1960; submitted to Nuclear Instruments and Methods.

FIGURE CAPTIONS

- Fig. 1. Block diagram of the coincidence-discriminator circuit.
- Fig. 2. Schematic diagram of the coincidence-discriminator circuit.
- Fig. 3. Schematic diagram of the five-channel driver unit.
- Fig. 4. Alternate coincidence circuit.
- Fig. 5. Coincidence discriminator unit.



MU - 21847

Fig. 1








Fig. 3



MU-21848

Fig. 4



ZN-2710

Fig. 5

This report was prepared as an account of Government sponsored work. Neither the United States, nor the Commission, nor any person acting on behalf of the Commission:

- A. Makes any warranty or representation, expressed or implied, with respect to the accuracy, completeness, or usefulness of the information contained in this report, or that the use of any information, apparatus, method, or process disclosed in this report may not infringe privately owned rights; or
- B. Assumes any liabilities with respect to the use of, or for damages resulting from the use of any information, apparatus, method, or process disclosed in this report.

As used in the above, "person acting on behalf of the Commission" includes any employee or contractor of the Commission, or employee of such contractor, to the extent that such employee or contractor of the Commission, or employee of such contractor prepares, disseminates, or provides access to, any information pursuant to his employment or contract with the Commission, or his employment with such contractor.