

**A superconducting pressure test-station for
material property tests at 4.2 K¹**

**J.M.van Oort,
Lawrence Berkeley Laboratory,
Accelerator and Fusion Research Division,
Superconducting Magnet Group,
1, Cyclotron Road, Berkeley, CA 94720**

ABSTRACT

A material property test system for use in liquid helium utilizing two repelling superconducting solenoids is described. The system is optimized for three parameters : to obtain the highest possible pressure on a small test volume using an existing cryostat, keeping the system as simple as possible in operation, and finally minimizing the use of liquid helium. To eliminate the use of a large hydrolic pressure system with all the related problems of transferring the applied pressure to the sample while maintaining a low heat-leak, a design with two superconducting coils was chosen. The system is designed to generate a force of 260 kN at a coil current of 100 A, thus yielding a pressure of 210 MPa at a 40 mm diameter maximum radius.

¹ This work was supported by the Director, Office of Energy Research, Office of High Energy and Nuclear Physics, High Energy Physics Division, U.S. Department of Energy, under contract No. DE-AC03-76SF00098.